

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/01/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:08
INTERNAL STANDARD	VERIFIED	01:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:09
BLANK TEST	0.00	01:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:11
BLANK TEST	0.00	01:12

WFO 2118

