

SOLUTION CHANGE




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/02/2026

OPERATOR NAME:
BRADY A WOOD
Permit #: 52004
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:03
INTERNAL STANDARD	VERIFIED	23:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:03
BLANK TEST	0.00	23:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:04
BLANK TEST	0.00	23:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:06
BLANK TEST	0.00	23:06



WFO - Solution Change

BW^{#66} 7/8/26