

WFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/10/2026

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:39
INTERNAL STANDARD	VERIFIED	18:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:39
BLANK TEST	0.00	18:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:40
BLANK TEST	0.00	18:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:42
BLANK TEST	0.00	18:43

