

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/03/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:10
INTERNAL STANDARD	VERIFIED	23:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:10
BLANK TEST	0.00	23:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:11
BLANK TEST	0.00	23:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:13
BLANK TEST	0.00	23:13

WFO @ LER

