

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/08/2026

OPERATOR NAME:
BENJAMIN E PETRIE
Permit #: 52840
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:35
INTERNAL STANDARD	VERIFIED	18:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:36
BLANK TEST	0.00	18:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:37
BLANK TEST	0.00	18:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:38
BLANK TEST	0.00	18:39

