

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/19/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

WFO

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:28

