

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/21/2026

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26270

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:58
INTERNAL STANDARD	VERIFIED	05:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:58
BLANK TEST	0.00	05:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:59
BLANK TEST	0.00	06:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:01
BLANK TEST	0.00	06:01

