## SUPERVISOR TEST

---

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

---

Date: 10/28/2026

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26270

Nominal = 0.10 g/210L

| | | |
|---|---|---|
| BLANK TEST | 0.00 | 01:21 |
| INTERNAL STANDARD | VERIFIED | 01:21 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.09 | 01:22 |
| BLANK TEST | 0.00 | 01:22 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.09 | 01:23 |
| BLANK TEST | 0.00 | 01:24 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.10 | 01:24 |
| BLANK TEST | 0.00 | 01:25 |