

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/29/2026

OPERATOR NAME:
RAYMOND Y BOYKO
Permit #: 53011
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:17
INTERNAL STANDARD	VERIFIED	07:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	07:17
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:19
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:21

WFO
Weekly

