

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/30/2026

OPERATOR NAME:
BENJAMIN E PETRIE
Permit #: 52840
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:18
INTERNAL STANDARD	VERIFIED	18:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:18
BLANK TEST	0.00	18:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:20
BLANK TEST	0.00	18:21

