

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/16/2026

OPERATOR NAME:
BENJAMIN E PETRIE
Permit #: 52840
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:42
INTERNAL STANDARD	VERIFIED	17:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:42
BLANK TEST	0.00	17:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:43
BLANK TEST	0.00	17:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:45
BLANK TEST	0.00	17:45

