

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/24/2026

OPERATOR NAME:
BENJAMIN E PETRIE
Permit #: 52840
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:28
INTERNAL STANDARD	VERIFIED	18:28
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:29
BLANK TEST	0.00	18:29
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:30
BLANK TEST	0.00	18:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:31
BLANK TEST	0.00	18:32

