

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 03/18/2026

OPERATOR NAME:  
BENJAMIN E PETRIE  
Permit #: 52840  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:36
INTERNAL STANDARD	VERIFIED	18:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:36
BLANK TEST	0.00	18:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:37
BLANK TEST	0.00	18:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:38
BLANK TEST	0.00	18:39

