

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 03/11/2026

OPERATOR NAME: *WFO*  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:11
INTERNAL STANDARD	VERIFIED	18:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:12
BLANK TEST	0.00	18:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:13
BLANK TEST	0.00	18:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:14
BLANK TEST	0.00	18:15

