

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/08/2026

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:02
INTERNAL STANDARD	VERIFIED	18:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:03
BLANK TEST	0.00	18:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:04
BLANK TEST	0.00	18:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:05
BLANK TEST	0.00	18:06

