

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 03/04/2026

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:38
INTERNAL STANDARD	VERIFIED	23:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:40
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42

