

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/15/2026

OPERATOR NAME:
RAYMOND Y BOYKO
Permit #: 52011
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:32
INTERNAL STANDARD	VERIFIED	06:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:32
BLANK TEST	0.00	06:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:34
BLANK TEST	0.00	06:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:35

RFI Detected

WFO

