

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/12/2026

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:34
INTERNAL STANDARD	VERIFIED	05:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:34
BLANK TEST	0.00	05:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:35
BLANK TEST	0.00	05:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:37
BLANK TEST	0.00	05:37

