

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/26/2026

OPERATOR NAME:
RAYMOND Y BOYKO
Permit #: 53011
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:26
INTERNAL STANDARD	VERIFIED	09:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	09:26
BLANK TEST	0.00	09:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	09:28
BLANK TEST	0.00	09:28
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	09:29
BLANK TEST	0.00	09:30

WFO
Weekly

