

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/05/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:40
INTERNAL STANDARD	VERIFIED	19:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:41
BLANK TEST	0.00	19:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:42
BLANK TEST	0.00	19:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:43
BLANK TEST	0.00	19:44

