

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/05/2026

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:18
INTERNAL STANDARD	VERIFIED	22:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:18
BLANK TEST	0.00	22:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:19
BLANK TEST	0.00	22:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:20
BLANK TEST	0.00	22:21

