SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/23/2026

OPERATOR NAME:
RAYMOND Y BOYKO
Permit #: 53011
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:02
INTERNAL STANDARD	VERIFIED	06:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:03
BLANK TEST	0.00	06:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:04
BLANK TEST	0.00	06:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:05
BLANK TEST	0.00	06:06

weekly

WFO