

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 01/21/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:13
INTERNAL STANDARD	VERIFIED	18:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:13
BLANK TEST	0.00	18:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:14
BLANK TEST	0.00	18:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:15
BLANK TEST	0.00	18:16

