

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 01/15/2026

OPERATOR NAME:
RAYMOND Y BOYKO
Permit #: 53011
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:14
INTERNAL STANDARD	VERIFIED	14:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:14
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:15
BLANK TEST	0.00	14:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17

WFO
Weekly
1/15/26

