

WFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 01/28/2026

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:56
INTERNAL STANDARD	VERIFIED	05:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:56
BLANK TEST	0.00	05:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:57
BLANK TEST	0.00	05:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:59
BLANK TEST	0.00	05:59

