

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 01/08/2026

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:49
INTERNAL STANDARD	VERIFIED	05:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:50
BLANK TEST	0.00	05:51
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:51
BLANK TEST	0.00	05:52
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:52
BLANK TEST	0.00	05:53

