

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/11/2026

WFB

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:43
INTERNAL STANDARD	VERIFIED	18:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:43
BLANK TEST	0.00	18:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:44
BLANK TEST	0.00	18:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:46
BLANK TEST	0.00	18:46

