

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 02/04/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:35
INTERNAL STANDARD	VERIFIED	01:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:35
BLANK TEST	0.00	01:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:37
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:38
BLANK TEST	0.00	01:39

