

# SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 04/22/2026

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 26050

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 18:01 |
| INTERNAL STANDARD       | VERIFIED | 18:01 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 18:01 |
| BLANK TEST              | 0.00     | 18:02 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 18:02 |
| BLANK TEST              | 0.00     | 18:03 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 18:03 |
| BLANK TEST              | 0.00     | 18:04 |

