

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/25/2026

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:08
INTERNAL STANDARD	VERIFIED	18:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:08
BLANK TEST	0.00	18:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:09
BLANK TEST	0.00	18:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:11
BLANK TEST	0.00	18:11

